

REPORTE TÉCNICO DECRETO SUPREMO N°38/11 DEL MINISTERIO DEL MEDIO AMBIENTE

Norma de Emisión de Ruidos Generados por Fuentes que Indica


FICHA DE EVALUACIÓN DE NIVELES DE RUIDO					
IDENTIFICACIÓN DE LA FUENTE EMISORA DE RUIDO					
Nombre Fuente Emisora	SOCIEDAD GASTRONÓMICA CHINA XUAN LTDA.				
Nombre o Razón Social	SOCIEDAD GASTRONÓMICA CHINA XUAN LTDA.				
RUT	76509801-7				
Dirección	RTA Q 45 CAMINO A ANTUCO KM. 3 0	Comuna	Los Ángeles		
Tipo de Fuente	Actividad Comercial	Subtipo Fuente	Restaurant		
RESUMEN DE EVALUACIÓN					
Punto de medición	NPC [dBA]	Zona D.S. 38/11 MMA	Período (Diurno/Nocturno)	Límite [dBA]	Estado (Supera/No supera)
1 - 1	59	Zona Rural	Nocturno	50	Supera en 9 dBA
1 - 2	57	Zona Rural	Nocturno	50	Supera en 7 dBA
OBSERVACIONES DEL PROCESO DE MEDICIÓN					
Sin observaciones					

IDENTIFICACIÓN DEL INFORME TÉCNICO	
Fecha de emisión	18/03/2024
Nombre encargado medición	Hugo Ramirez
Institución o empresa	SMA

REPORTE TÉCNICO DECRETO SUPREMO N°38/11 DEL MINISTERIO DEL MEDIO AMBIENTE

Norma de Emisión de Ruidos Generados por Fuentes que Indica

FICHA DE GEOREFERENCIACIÓN DE MEDICIÓN DE RUIDO



LEYENDA DE CROQUIS O IMAGEN UTILIZADA

DATUM	WGS84	Huso	18S	
Fuente	Símbolo	Nombre	Coordenadas	
	FE	SOCIEDAD GASTRONÓMICA CHINA XUAN LTDA.	N	5850678
			E	739206

RECEPTORES

Símbolo	Nombre	Coordenadas	
R1	1	N	5850759
		E	739178
RFA	Ruido de Fondo A	N	5853271
		E	741290
RFB	Ruido de Fondo B	N	5850828
		E	739225

REPORTE TÉCNICO DECRETO SUPREMO N°38/11 DEL MINISTERIO DEL MEDIO AMBIENTE

Norma de Emisión de Ruidos Generados por Fuentes que Indica

FICHA DE INFORMACIÓN DE MEDICIÓN DE RUIDO

IDENTIFICACIÓN DE RECEPTORES

Nombre o Razón Social	1		
Dirección	Ruta Q-45 S/N	Comuna	Los Ángeles
Zona IPT	Rural	Homologación	Zona Rural
Descripción del Receptor	Vivienda consolidada de hormigón.		

REPORTE TÉCNICO DECRETO SUPREMO N°38/11 DEL MINISTERIO DEL MEDIO AMBIENTE

Norma de Emisión de Ruidos Generados por Fuentes que Indica

FICHA DE MEDICIÓN DE NIVELES DE RUIDO

Nombre o Razón Social Receptor	1	Número Medición	1
Fecha de medición	20/01/2024	Perido de medición	Nocturno
Hora inicio de medición	23:00	Hora término de medición	23:55
Condición de medición	Externa	Condición ventana	No Aplica
Descripción lugar de medición	Terraza vivienda con pared colindante a la Fuente Emisora		
Identificación del ruido de fondo	Fronda de árboles, crujido de insectos y paso esporádico de vehículos con tubo de escape abierto los cuales son filtrados.		

INSTRUMENTAL DE MEDICIÓN

CARACTERÍSTICA	SONÓMETRO	CALIBRADOR ACÚSTICO
Marca	CIRRUS	CIRRUS
Modelo	162B	CR154
Nº de serie	G304081	99724
Fecha certificado de calibración	21/12/2022	21/12/2022
Código certificado de calibración	185057	185034

FICHA DE EVALUACIÓN DE NIVELES DE RUIDO

Descriptor	MEDICIÓN 1			MEDICIÓN 2			MEDICIÓN 3		
	1	2	3	4	5	6	7	8	9
NPSeq	59.2	59.6	58.1	-	-	-	-	-	-
NPSmáx	61.5	62.9	60.6	-	-	-	-	-	-
NPSmin	49.5	54.5	55.1	-	-	-	-	-	-

REGISTRO RUIDO DE FONDO

Afecta medición	Si	Fecha	20/01/2024	Hora	20:30				
	5'	10'	15'	20'	25'	30'	Medición realizada en punto receptor		

COORDENADAS DE MEDICIÓN DE RUIDO DE FONDO

DATUM	WGS84	HUSO	18S
Coordenada Norte	5853271	Coordenada Este	741290

RESULTADO DE MEDICIÓN

RUIDO DE FONDO	NPC
43	59

OBSERVACIONES

El ruido de fondo se efectuó en sector donde no fuera posible percibir a la fuente emisora

REPORTE TÉCNICO DECRETO SUPREMO N°38/11 DEL MINISTERIO DEL MEDIO AMBIENTE

Norma de Emisión de Ruidos Generados por Fuentes que Indica

FICHA DE MEDICIÓN DE NIVELES DE RUIDO

Nombre o Razón Social Receptor	1	Número Medición	2
Fecha de medición	01/03/2024	Perido de medición	Nocturno
Hora inicio de medición	21:31	Hora término de medición	22:27
Condición de medición	Externa	Condición ventana	No Aplica
Descripción lugar de medición	Terraza vivienda cuyo muro colinda con Fuente Emisora		
Identificación del ruido de fondo	Fronda de árboles, ladridos, trinar de aves y tránsito vehicular.		

INSTRUMENTAL DE MEDICIÓN

CARACTERÍSTICA	SONÓMETRO	CALIBRADOR ACÚSTICO
Marca	CIRRUS	CIRRUS
Modelo	162B	CR154
Nº de serie	G304081	99724
Fecha certificado de calibración	21/12/2022	21/12/2022
Código certificado de calibración	185057	185034

FICHA DE EVALUACIÓN DE NIVELES DE RUIDO

Descriptor	MEDICIÓN 1			MEDICIÓN 2			MEDICIÓN 3		
	1	2	3	4	5	6	7	8	9
NPSeq	56.7	57.2	57.6	-	-	-	-	-	-
NPSmáx	59.8	59.6	59.1	-	-	-	-	-	-
NPSmin	54.2	54.1	53.8	-	-	-	-	-	-

REGISTRO RUIDO DE FONDO

Afecta medición	Si	Fecha	01/03/2024	Hora	18:35				
	5'	10'	15'	20'	25'	30'	Medición realizada en punto receptor		
NPSeq	44.7	44	-	-	-	-	No		

COORDENADAS DE MEDICIÓN DE RUIDO DE FONDO

DATUM	WGS84			HUSO	18S
Coordenada Norte	5850828			Coordenada Este	739225

RESULTADO DE MEDICIÓN

RUIDO DE FONDO	NPC
44	57

OBSERVACIONES

Fuente emisora al momento de la medición de ruido de fondo no estaba en funcionamiento


CERTIFICATE OF CALIBRATION

ISSUED BY Cirrus Research plc

DATE OF ISSUE 21 December 2022 CERTIFICATE NUMBER 185034



10148



Cirrus Research plc
Acoustic House
Bridlington Road
Hunmanby
North Yorkshire
YO14 0PH
United Kingdom

Page 1 of 2

Approved signatory

C.Scott

Electronically signed:

A handwritten signature of "C.Scott" in black ink.

Sound Calibrator : IEC 60942:2003

Customer information

Name: Sociedad Acustical S.A. Address: Villaseca 21
Oficina 1104
Nunoa
Santiago Postcode:
Country: Chile

Instrument information

Manufacturer: Cirrus Research plc Notes:

Model: CR:514

Serial number: 99724

Class: 2

Pattern approval: Yes

Source of pattern approval: PTB-1.61-4028829

Test summary

Date of receipt: 21 December 2022

Date of calibration: 21 December 2022

As public evidence was available, from a testing organisation responsible for approving the results of pattern evaluation tests, to demonstrate that the model of sound calibrator fully conformed to the requirements for pattern evaluation described in Annex A of IEC 60942:2003, the sound calibrator tested is considered to conform to all the Class 2 requirements of IEC 60942:2003.

Notes

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. UKAS is one of the signatories to the Multilateral Agreement of the European co-operation for Accreditation (EA) for the mutual recognition of calibration certificates issued by accredited laboratories. The United Kingdom Accreditation Service (UKAS) is one of the signatories to the International Laboratory Accreditation Cooperation (ILAC) Arrangement for the mutual recognition of calibration certificates. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. The results within this certificate relate only to the items calibrated. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability

CERTIFICATE OF CALIBRATION

Certificate Number:

185034

Page 2 of 2

Environmental conditions

The following conditions were recorded at the time of the test:

Pressure: 99.20 kPa
Temperature: 23.3 °C
Humidity: 36.8 %

Test equipment

Equipment	Manufacturer	Model	Serial number
Calibrator	Brüel and Kjaer	4231	2564324
Microphone	G.R.A.S	40AP	173198
Distortion Meter	Keithley	2015	0761605
Multimeter	TTi	1908	509541

Calibration procedure

The sound calibrator under test has been calibrated to the published data as described in the operating manual and in the half-inch configuration.

The procedures and techniques used are as described in IEC 60942:2003 Annex B – Periodic Tests. Firstly, five determinations of the sound pressure level were made using the Reference calibrator. Then, five determinations of the sound pressure level, frequency and total distortion were made using the calibrator under test.

Measurements

Reference calibrator SPL: 94.01 dB


Configuration	Output Level (dB)		Frequency (Hz)	Total distortion (%)
94 dB	Measured	94.00	1000.3	0.5
	Deviation	+0.00	+0.3	
Tolerance		±0.75	±20.0	≤ 4.0
Uncertainty		±0.09	±0.1	±0.21

CERTIFICATE OF CALIBRATION



ISSUED BY Cirrus Research plc

DATE OF ISSUE 21 December 2022 CERTIFICATE NUMBER 185057



Cirrus Research plc
Acoustic House
Bridlington Road
Hunmanby
North Yorkshire
YO14 0PH
United Kingdom

Page 1 of 15

Approved signatory

C.Scott

Electronically signed:

Sound level meter : IEC 61672-3:2013

Customer information

Name: Sociedad Acustical S.A. Address: Villaseca 21 Postcode:
Oficina 1104
Nunoa
Santiago Country: Chile

Instrument information

Manufacturer:	Cirrus Research plc	Notes:
Model:	CR:162B	
Serial number:	G304081	
Class:	2	
Firmware version:	V5.8.3251	

Test summary

Date of receipt: 21 December 2022 Date of calibration: 21 December 2022

Periodic tests were performed in accordance with procedures from IEC 61672-3:2013.

The sound level meter submitted for testing successfully completed the class 2 periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed.

However, no general statement or conclusion can be made about conformance of the sound level meter to the full specifications of IEC 61672-1:2013 because (a) evidence was not publicly available, from an independent testing organisation responsible for pattern approvals, to determine that the model of sound level meter fully conformed to the class 2 specifications in IEC 61672-1:2013 or correction data for acoustical test of frequency weighting were not provided in the Instruction Manual and (b) because the periodic tests of IEC 61672-3:2013 cover only a limited subset of the specifications in IEC 61672-1:2013.

Notes

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Certificate Number:

185057

Page 2 of 15

Environmental conditions

The following conditions were recorded at the time of the test:

Before	Pressure: 99.13 kPa	Temperature: 23.6 °C	Humidity: 40.0 %
After	Pressure: 99.12 kPa	Temperature: 23.4 °C	Humidity: 40.8 %

Test equipment

Equipment	Manufacturer	Model	Serial number
Signal Generator	KEYSIGHT	33511B	MY58001681
Attenuator	Cirrus Research	ZE:952	78701
Multi-frequency Calibrator	B&K	4226	3309971

Additional instrument information

Instruction manual: Part B Manual
Reference level range: Single range
Pattern approval: No
Source of pattern approval: -

Preamplifier

Manufacturer: Cirrus Research plc
Model: MV:200F
Serial number: 12666F

Microphone

Manufacturer: Cirrus Research plc
Model: MK:216
Serial number: 414391D

Test results summary

Test	Result
Self-generated noise	Complies
Long-term stability	Complies
Acoustic frequency weightings	Complies
Electrical frequency weightings	Complies
Weightings at 1 kHz	Complies
Linearity	Complies
Toneburst response	Complies
C-weighted peak	Complies
Overload	Complies
High-level stability	Complies

Acoustic Calibrator

Manufacturer: Cirrus Research plc
Model: CR:514
Serial number: 99724

Calibration

Calibration check frequency: 1000 Hz
Calibrator's certificate ref: 185034
Level before adjustment: 93.80 dB(A)
Level after adjustment: 93.70 dB(A)

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Certificate Number:
185057

Page 3 of 15

Laboratory uncertainties

Requirement	Value (dB)
Frequency weightings at 1 kHz	0.1
Acoustic freq. weighting 1 kHz	0.21
Acoustic freq. weighting 125 Hz	0.21
Acoustic freq. weighting 8 kHz	0.23
Electrical freq. weighting 1 kHz	0.15
Electrical freq. weighting 125 Hz	0.15
Electrical freq. weighting 16 kHz	0.17
Electrical freq. weighting 2 kHz	0.15
Electrical freq. weighting 250 Hz	0.15
Electrical freq. weighting 4 kHz	0.15
Electrical freq. weighting 500 Hz	0.15
Electrical freq. weighting 63 Hz	0.15
Electrical freq. weighting 8 kHz	0.15
High level stability	0.1
Level linearity	0.15
Long-term stability	0.1
Overload indication	0.1
Peak C sound level	0.12
Time weightings at 1 kHz	0.1
Toneburst response	0.1

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Certificate Number:

185057

Page 4 of 15

Section 11: Self-generated noise

This test is for indication only and a failure does not, on its own, mean the sound level meter does not comply to the requirements of IEC 61672-1.

Selected range: Single range

		Acoustic	Electrical		
		LAeq	LAeq	LCeq	LZeq
Level	Measured	15.10	Under-range	14.20	26.60
	Highest anticipated	19.00	Under-range	24.00	40.00

All values in the table are given in decibels.

Section 14: Frequency and time weightings at 1 kHz

Result: Passed

Selected range: Single range

Frequency weightings at 1 kHz

LAF	LCF	LZF	LCF - LAF	LZF - LAF	Acceptance limit	Uncertainty
94.00	94.00	94.00	0.00	0.00	±0.2	±0.12

All values in the table are given in decibels.

Time weightings at 1 kHz

LAF	LAS	LAeq	LAS - LAF	LAeq - LAF	Acceptance limit	Uncertainty
94.00	94.00	94.00	0.00	0.00	±0.1	±0.12

All values in the table are given in decibels.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Section 12: Acoustical signal tests of a frequency weighting

Tested range: Single range

Source of multi-frequency calibrator correction: Multi-Cal Internal Corrections

Source of instrument case correction: Part B Manual V20160609

Source of windshield correction: Part B Manual V20160609

Frequency	L _{Ceq} 1	L _{Ceq} 2	L _{Ceq} 3	Average L _{Ceq}	Corr. A*	Corr. B*	Corr. C*	Corr. D*	Corrected L _{Ceq}	Rel. freq-weighting	Design goal	Deviation	Total uncertainty	Acceptance limit
125 Hz	93.80	93.80	93.80	93.80	0.00	0.00	0.00	0.00	93.80	-0.16	-0.2	0.0	0.46	±1.5
1000 Hz	93.80	93.80	93.80	93.80	-0.03	0.00	-0.03	0.22	93.96				0.46	
10000 Hz	89.10	89.10	89.10	89.10	0.03	0.00	-0.08	0.04	89.09	-4.87	-3.0	-1.9	0.62	±5.0

All values in the table are given in decibels unless otherwise stated.

end
correction for the difference between the multi-frequency calibrator input signal level at a test frequency and at 1 kHz

multi-frequency calibrator correction

correction for the effect of the instrument case

correction for the effect of the windscreens

Information on the uncertainty of measurement of the adjustment data given in the instruction manual or obtained from the manufacturer or supplier of the sound level meter published in the instruction manual or made available by the manufacturer or supplier. The uncertainty of measurement of the adjustment data was assumed to be empirically the maximum-permitted uncertainty given in IEC 62585 for the purpose of the periodic test.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Section 13: Electrical signal tests of frequency weightings : A-Weighting

Selected range: Single range

Source of microphone free-field correction: Actuator Test (Cert: 185032)

Source of instrument case correction: Part B Manual V20160609

Source of windshield correction: Part B Manual V20160609

Frequency	Equivalent input level	L _{Aeq}	Deviation from 1 kHz	Correction A*	Correction B*	Correction C*	Corrected deviation	Total uncertainty	Acceptance limit
63 Hz	120.20	94.40	0.40	0.25	0.00	0.00	0.7	0.35	±2.0
125 Hz	110.10	94.20	0.20	0.16	0.00	0.00	0.4	0.35	±1.5
250 Hz	102.60	94.20	0.20	0.10	0.02	0.03	0.4	0.35	±1.5
500 Hz	97.20	94.10	0.10	0.07	0.01	0.07	0.3	0.35	±1.5
1000 Hz	94.00	94.00	0.00	-0.03	-0.03	0.22	0.2	0.35	±1.0
2000 Hz	92.80	93.90	-0.10	-0.18	-0.01	0.59	0.3	0.35	±2.0
4000 Hz	93.00	93.70	-0.30	-0.62	-0.03	0.13	-0.8	0.35	±3.0
8000 Hz	95.10	93.50	-0.50	-0.90	-0.08	0.04	-1.4	0.49	±5.0

All values are given in decibels unless otherwise stated.

Legend

Microphone free-field correction

Correction for the effect of the instrument case

Correction for the effect of the windscreen

Information on the uncertainty of measurement, required by IEC 61672-3:2013, for the correction data given in the Instruction Manual or obtained from the manufacturer or supplier of the sound level meter, or the manufacturer of the microphone, was provided in the Instruction Manual or made available by the manufacturer or supplier of the sound level meter. The uncertainty of measurement of the correction data was therefore assumed to be the maximum-permitted uncertainty given in IEC 62585 for the corresponding free-field correction data and for a coverage probability of 95 %.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

13: Electrical signal tests of frequency weightings : C-Weighting

ected range: Single range

ce of microphone free-field correction: Actuator Test (Cert: 185032)

ce of instrument case correction: Part B Manual V20160609

ce of windshield correction: Part B Manual V20160609

Frequency	Equivalent input level	L _{eq}	Deviation from 1 kHz	Correction A*	Correction B*	Correction C*	Corrected deviation	Total uncertainty	Acceptance limit
63 Hz	94.80	94.00	0.00	0.25	0.00	0.00	0.3	0.35	±2.0
125 Hz	94.20	94.00	0.00	0.16	0.00	0.00	0.2	0.35	±1.5
250 Hz	94.00	94.00	0.00	0.10	0.02	0.03	0.2	0.35	±1.5
500 Hz	94.00	94.00	0.00	0.07	0.01	0.07	0.2	0.35	±1.5
1000 Hz	94.00	94.00	0.00	-0.03	-0.03	0.22	0.2	0.35	±1.0
2000 Hz	94.20	93.90	-0.10	-0.18	-0.01	0.59	0.3	0.35	±2.0
4000 Hz	94.80	93.80	-0.20	-0.62	-0.03	0.13	-0.7	0.35	±3.0
8000 Hz	97.00	93.60	-0.40	-0.90	-0.08	0.04	-1.3	0.49	±5.0

All values are given in decibels unless otherwise stated.

Legend
Microphone free-field correction
Correction for the effect of the instrument case
Correction for the effect of the windscreens

Information on the uncertainty of measurement, required by IEC 61672-3:2013, for the correction data given in the Instruction Manual or obtained from the manufacturer or supplier of the sound level meter, or the manufacturer of the microphone, was provided in the Instruction Manual or made available by the manufacturer or supplier of the sound meter. The uncertainty of measurement of the correction data was therefore assumed to be the maximum-permitted uncertainty given in IEC 62585 for the corresponding free-field correction data and for a coverage probability of 95 %.

Certificate Number: 185057
Page 7 of 15

Result: Passed

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Section 13: Electrical signal tests of frequency weightings : Z-Weighting

Selected range: Single range

Source of microphone free-field correction: Actuator Test (Cert: 185032)

Source of instrument case correction: Part B Manual V20160609

Source of windshield correction: Part B Manual V20160609

Frequency	Equivalent input level	L _{eq}	Deviation from 1 kHz	Correction A*	Correction B*	Correction C*	Corrected deviation	Total uncertainty	Acceptance limit
63 Hz	94.00	94.20	0.20	0.25	0.00	0.00	0.5	0.35	±2.0
125 Hz	94.00	94.10	0.10	0.16	0.00	0.00	0.3	0.35	±1.5
250 Hz	94.00	94.10	0.10	0.10	0.02	0.03	0.3	0.35	±1.5
500 Hz	94.00	94.00	0.00	0.07	0.01	0.07	0.2	0.35	±1.5
1000 Hz	94.00	94.00	0.00	0.00	-0.03	0.22	0.2	0.35	±1.0
2000 Hz	94.00	94.00	0.00	-0.18	-0.01	0.59	0.4	0.35	±2.0
4000 Hz	94.00	94.00	0.00	-0.62	-0.03	0.13	-0.5	0.35	±3.0
8000 Hz	94.00	93.90	-0.10	-0.90	-0.08	0.04	-1.0	0.49	±5.0

Legend

Microphone free-field correction

Correction for the effect of the instrument case

Correction for the effect of the windscreen

All values are given in decibels unless otherwise stated.

Information on the uncertainty of measurement, required by IEC 61672-3:2013, for the correction data given in the Instruction Manual or obtained from the manufacturer or supplier of the sound level meter, or the manufacturer of the microphone, was provided in the Instruction Manual or made available by the manufacturer or supplier of the sound level meter. The uncertainty of measurement of the correction data was therefore assumed to be the maximum-permitted uncertainty given in IEC 62585 for the corresponding free-field correction data and for a coverage probability of 95 %.

Certificate Number: 185057
Page 8 of 15

Result: Passed


CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Section 13: Electrical signal tests of frequency weightings : Continued

Certificate Number:	185057
Page 9 of 15	

Result: Passed



CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Certificate Number:

185057

Page 10 of 15

Result: Passed

Section 16: Level linearity on the reference level range

Selected range: Single range **Acceptance limit** ± 1.1 dB **Uncertainty:** 0.15 dB

Frequency: 8000 Hz **Linear operating range at 8000 kHz:** 24 - 137 dB

Reference level: 94.00 dB **Ref. sig. gen. voltage:** 0.219 Vpp **Ref.attenuator gain:** 0 dB

Input level	LAeq	Deviation
138.00	138.00	0.0
137.00	137.00	0.0
136.00	136.00	0.0
135.00	135.00	0.0
134.00	134.00	0.0
133.00	133.00	0.0
132.00	132.00	0.0
129.00	129.00	0.0
124.00	124.00	0.0
119.00	119.00	0.0
114.00	114.00	0.0
109.00	109.00	0.0
104.00	104.00	0.0
99.00	99.00	0.0
94.00	94.00	0.0
89.00	89.00	0.0
84.00	84.10	0.1
79.00	79.10	0.1
74.00	74.10	0.1
69.00	69.10	0.1
64.00	64.10	0.1
59.00	59.10	0.1
54.00	54.00	0.0
49.00	49.00	0.0
44.00	44.00	0.0
39.00	39.00	0.0
34.00	34.00	0.0
29.00	29.00	0.0
28.00	28.00	0.0
27.00	27.00	0.0
26.00	26.00	0.0
25.00	25.00	0.0

All values in the table are given in decibels.


CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Certificate Number:
185057

Page 11 of 15

Section 16: Level linearity on the reference level range



CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Section 18: Toneburst response

Selected range: Single range
Frequency: 4000 Hz

t: Steady level type: LAF Burst level type: LAFMax

Input level: 137.00 dB
Uncertainty: 0.12 dB

Burst duration	Steady level	Burst level	Measured difference	Target difference	Deviation	Acceptance limit	Overload
200 ms	137.00	136.00	-1.00	-1.0	0.0	±1.0	No
2 ms	137.00	118.90	-18.10	-18.0	-0.1	+1.0; -2.5	No
0.25 ms	137.00	109.90	-27.10	-27.0	-0.1	+1.5; -5.0	No

All values in the table are given in decibels unless otherwise stated.

w: Steady level type: LAS Burst level type: LASMax

Burst duration	Steady level	Burst level	Measured difference	Target difference	Deviation	Acceptance limit	Overload
200 ms	137.00	129.60	-7.40	-7.4	0.0	±1.0	No
2 ms	137.00	110.00	-27.00	-27.0	0.0	+1.0; -5.0	No

All values in the table are given in decibels unless otherwise stated.

..: Steady level type: LAeq Burst level type: LAE

Burst duration	Steady level	Burst level	Measured difference	Target difference	Deviation	Acceptance limit	Overload
200 ms	137.00	130.00	-7.00	-7.0	0.0	±1.0	No
2 ms	137.00	110.00	-27.00	-27.0	0.0	+1.0; -2.5	No
0.25 ms	137.00	100.90	-36.10	-36.0	-0.1	+1.5; -5.0	No

All values in the table are given in decibels unless otherwise stated.

Certificate Number: 185057
Page 12 of 15

Result: Passed

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Test 19: C-weighted peak sound level

Tested range: Single range Uncertainty: 0.12 dB

Frequency	Test signal	Reference LCeq	LCPeak	Measured difference	Target difference	Deviation	Acceptance limit	Overload
8000 Hz	Full sinewave	130.00	133.50	3.50	3.4	0.1	±3.0	No
500 Hz	Positive half-sinewave	134.00	136.30	2.30	2.4	-0.1	±2.0	No
500 Hz	Negative half-sinewave	134.00	136.30	2.30	2.4	-0.1	±2.0	No

All values in the table are given in decibels.

Certificate Number: 185057
Page 13 of 15

Result: Passed

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Certificate Number:

185057

Page 14 of 15

Section 15: Long-term stability

Result: Passed

Frequency: 1000 Hz

Selected range: Single range

Reference level: 94.00 dB

Reference input level sig. gen. voltage: 0.181 Vpp

Reference input level attenuator gain: 0 dB

	Before	After	Difference	Acceptance limit	Uncertainty
L _{Aeq}	94.00	94.00	0.0	0.3	0.10

All values in the table are given in decibels.

Section 21: High-level stability

Result: Passed

Frequency: 1000 Hz

Selected range: Single range

Reference level: 138.00 dB

	Before	After	Difference	Acceptance limit	Uncertainty
L _{Aeq}	138.00	138.00	0.0	0.3	0.10

All values in the table are given in decibels.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 10148

Certificate Number:
185057

Page 15 of 15

Section 20: Overload indication

Result: Passed

Frequency: 4000 Hz
Uncertainty: ± 0.13 dB
Selected range: Single range
Start input level: 139.0 dB
Start input level sig. gen. voltage: 9.474 Vpp
Start input level attenuator gain: 10 dB

Relative input level

Positive half-sinewave	Negative half-sinewave	Difference	Acceptance limit
140.7	140.6	-0.1	1.5

All values in the table are given in decibels unless otherwise stated.

Latching: It was verified that the overload indicator latches as specified in IEC 61672-1.


CERTIFICATE OF CALIBRATION

ISSUED BY

Cirrus Research plc

DATE OF ISSUE

21 December 2022 CERTIFICATE NUMBER 185032



Cirrus Research plc
Acoustic House
Bridlington Road
Hunmanby
North Yorkshire
YO14 0PH
United Kingdom

Page 1 of 2

Test engineer:
C.Scott
Electronically signed:

Microphone

Microphone capsule

Manufacturer: Cirrus Research plc

Model: MK:216

Serial Number: 414391D

Calibration procedure

Date of calibration: 21 December 2022

Open circuit: 55.5 mV/Pa

Sensitivity at 1 kHz: -25.1 dB rel 1 V/Pa

The microphone capsule detailed above has been calibrated to the published data as described in the operating manual of the associated sound level meter (where applicable).

The frequency response was measured using an electrostatic actuator in accordance with BS EN 61094-6:2005 with the free-field response derived via standard correction data traceable to a National Measurement Institute.

The absolute sensitivity at 1 kHz was measured using an acoustic calibrator conforming to IEC 60942:2003 Class 1.

Environmental conditions

Pressure: 99.19 kPa

Temperature: 23.5 °C

Humidity: 37.6 %

CERTIFICATE OF CALIBRATION

Certificate Number:

185032

Page 2 of 2

Free-Field Frequency Response : Tabular

Frequency (Hz)	Free-Field Sensitivity (dB rel 1 kHz)	Actuator Response (dB)
63	0.25	-0.39
80	0.23	-0.24
100	0.13	-0.14
125	0.16	-0.06
160	0.13	0.01
200	0.11	0.06
250	0.10	0.09
315	0.09	0.11
400	0.09	0.12
500	0.07	0.11
630	0.05	0.10
800	0.03	0.08
1 000	0.00	0.02
1 250	-0.03	-0.06
1 600	-0.10	-0.19
2 000	-0.18	-0.38
2 500	-0.32	-0.66
3 150	-0.41	-1.06
4 000	-0.62	-1.65
5 000	-0.78	-2.35
6 300	-0.97	-3.22
8 000	-0.90	-4.27
10 000	-0.56	-5.53
12 500	-0.06	-6.94
16 000	-0.21	-8.72
20 000	-2.32	-12.40

Free-Field Frequency Response : Graphical

